



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/764,148	01/23/2004	Hanson S. Gifford	022128-000510US	8153
65689	7590	08/16/2007	EXAMINER	
CIERRA, INC. &			JOHNSON III, HENRY M	
TOWNSEND & TOWNSEND AND CREW, LLP				
TWO EMBARCADERO CENTER			ART UNIT	PAPER NUMBER
EIGHTH FLOOR				3739
SAN FRANCISCO, CA 94111-3834				
			MAIL DATE	DELIVERY MODE
			08/16/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No.	Applicant(s)
	10/764,148	GIFFORD ET AL.
	Examiner	Art Unit
	Henry M. Johnson, III	3739

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) Responsive to communication(s) filed on 10 July 2007.
- 2a) This action is FINAL. 2b) This action is non-final.
- 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) Claim(s) 1,2,4,5,16,17,19-21 and 36 is/are pending in the application.
 - 4a) Of the above claim(s) 6-15,35 and 37-39 is/are withdrawn from consideration.
- 5) Claim(s) 25,27-34,40 and 41 is/are allowed.
- 6) Claim(s) 1,2,4,5,16,17,19-21 and 36 is/are rejected.
- 7) Claim(s) _____ is/are objected to.
- 8) Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) The specification is objected to by the Examiner.
- 10) The drawing(s) filed on 14 July 2005 is/are: a) accepted or b) objected to by the Examiner.

Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).

Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) All b) Some * c) None of:
 1. Certified copies of the priority documents have been received.
 2. Certified copies of the priority documents have been received in Application No. _____.
 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) Notice of References Cited (PTO-892)
- 2) Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
- 5) Notice of Informal Patent Application
- 6) Other: _____

Response to Arguments

Applicant's arguments filed July 10, 2007, have been fully considered but they are not persuasive. The examiner's assertion that display of the focus of the energy would be intuitive is substantiated by the inclusion of a reference to Bucholz et al. in the following rejections.

Applicant's Terminal Disclaimer is not effective, as it is not signed by the attorney of record.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Claims 1, 2, 4, 5, 16-17, 19-21 and 36 are rejected under 35 U.S.C. 103(a) as being unpatentable over WO 01/82778 to Focus Surgery, Inc. in view of U.S. Patent 6,236,875 to Bucholz et al. WO 01/82778 teaches an ablation system with visualization using transducers to transmit high intensity focused to a tissue target (abstract). Ultrasound transducers (Fig. 2, # 30) are disclosed in a catheter with a proximal and distal end (page 4, lines 10-16). The

transducers are used for both HIFU and imaging and can be separately controlled (page 5, line 15-20). The transducers may be phased in a manner to vary the focus of the array (page 5, line 32) thus providing a variable depth and steerable beam (page 8, lines 5-12). This inherently requires a controller and the steerable capability is interpreted as two-dimensional, thus capable of a linear ablation. A display unit is also disclosed, however, the display of the focal point of the HIFU is not disclosed. Bucholz et al. teach a surgical navigation system with positional and imaging capabilities. A processor may modify a display to indicate the position of the instrument or the **instrument's focal point**, such as by positioning a cursor (marker), with respect to the body elements (Col. 10, lines 55-58). It would have been obvious to one skilled in the art to include the display of the focal point of the surgical instrument as taught by Bucholz et al. in the invention of Focus Surgery, Inc to clearly indicate the target of the instrument as clearly taught by Bucholz et al.

Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 1, 4-5, 16 and 19-20 are provisionally rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 1, 3-5 of copending Application No. 11/494387. Although the conflicting claims are not identical, they are not patentably distinct from each other because they are an obvious change in scope based on intended use.

This is a provisional obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

Allowable Subject Matter

Claims 25, 27-34 and 40-41 are allowed.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).


A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Art Unit: 3739

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Henry M. Johnson, III whose telephone number is (571) 272-4768. The examiner can normally be reached on Monday through Friday from 6:00 AM to 3:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C. Dvorak can be reached on (571) 272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


Henry M. Johnson, III
Primary Examiner
Art Unit 3739